

UDEM Adriatic d.o.o. Radnička cesta 54/R3 10000 Zagreb, CROATIA

2024/03/22

YURIM MEDICAL CO.,LTD 43, Jeongjung 3-gil, Osong-eup, Heungdeok-gu, Cheongiu-si. Chungcheongbuk-do, Republic of Korea

**NOTIFIED BODY CONFIRMATION LETTER** 

Reference: 2024.MDR.1093-22-2.NBCL.0105

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that, UDEM ADRIATIC D.O.O., a Notified Body (NB) designated under Regulation (EU) 2017/745 (MDR) and identified by the number 2696 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR (on the date of 2022/07/13) and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR (on the date of 2022/07/13) with the following manufacturer:

YURIM MEDICAL CO.,LTD 43, Jeongjung 3-gil, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, Republic of Korea SRN Number (if available): KR-MF-000009011

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which UDEM Adriatic d.o.o. is also responsible for appropriate surveillance of the corresponding devices under Directive 93/42/EEC (MDD). Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but UDEM Adriatic d.o.o. has not yet taken the responsibility for appropriate surveillance of the corresponding devices under MDD.

UDEM Adriatic d.o.o.

Address: Radnička cesta 54/R3, Green Gold Centar, 10000 Zagreb

Phone: +385 (1) 4819 601 • Fax: +385 (1) 4819 434

E-mail: info@udemadriatic.com

UDFRM.235/01-08.05.2023/07.04.2023 Page 1 / 4



In the case of devices covered by certificates issued under MDD that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 March 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120(3c) of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Wellestablished technologies (WET sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of UDEM Adriatic d.o.o.

Zekeriya AYTAÇ

General Manager

UDEM Adriatic d.o.o.

Address: Radnička cesta 54/R3, Green Gold Centar, 10000 Zagreb

Green Gold Centar IV. Kat Zagreb 9601 Fax: (1) 4819434

Phone: +385 (1) 4819 601 • Fax: +385 (1) 4819 434

E-mail: info@udemadriatic.com

UDEM Adriatic d.o.o
UDEM OIB: 5037079784

Padnicka Cesta 14/93 Green Gate Cestar IV. Kar Zagieb
Tok() adriadol For Distilation
Www.cestar driadiccom

UDFRM.235/01-08.05.2023/07.04.2023 Page 2 / 4



## Table 1: Devices covered by this letter and for which UDEM Adriatic d.o.o. is also responsible for appropriate surveillance of the corresponding devices under Directive 93/42/EEC (MDD)

Device name or Basic UDI- DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD device	MDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
N/A	N/A	N/A	N/A

## Table 2: Devices covered by this letter and for which UDEM Adriatic d.o.o. is <u>NOT</u> responsible for appropriate surveillance of the corresponding devices under Directive 93/42/EEC (MDD)

Device name or Basic UDI- DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD device	MDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Sterile Polydioxanone Suture with Needle	Class III	N/A	Certificate 1: Full Quality Assurance System Certificate No: M.2016.106.7195  UDEM Uluslararası Belgelendirme Denetim Egitim Merkezi San. ve Tic. A.Ş. (NB2292)  Certificate 2: EC Design-Examination Certificate No: M.2016.106.7195-1  UDEM Uluslararası Belgelendirme Denetim Egitim Merkezi San. ve Tic. A.Ş. (NB2292)
Sterile P(LA/CL) [Poly(L-Lactide-co-ε- Caprolactone)] Suture with Needle	Class III	N/A	Certificate 1: Full Quality Assurance System Certificate No: M.2019.106.11991  UDEM Uluslararası Belgelendirme Denetim Egitim Merkezi San. ve Tic. A.Ş. (NB2292)

**UDEM Adriatic d.o.o.** 

Address: Radnička cesta 54/R3, Green Gold Centar, 10000 Zagreb

Phone: +385 (1) 4819 601• Fax: +385 (1) 4819 434

E-mail: info@udemadriatic.com UDFRM.235/01-08.05.2023/07.04.2023 Page 3 / 4



Device name or Basic UDI- DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD device	MDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
			Certificate 2: EC Design-Examination Certificate No: M.2019.106.11991-1
			UDEM Uluslararası Belgelendirme Denetim Egitim Merkezi San. ve Tic. A.Ş. (NB2292)

## **Confirmation Letter Revision History**

Date	NB internal reference traceable to each version of the letter	Action
2024/03/22	2024.MDR.1093-22-2.NBCL.0105	Initial issue



UDEM Adriatic d.o.o.

Address: Radnička cesta 54/R3, Green Gold Centar, 10000 Zagreb

**Phone**: +385 (1) 4819 601• **Fax**: +385 (1) 4819 434

E-mail: info@udemadriatic.com UDFRM.235/01-08.05.2023/07.04.2023 Page 4 / 4